

Stethoscope Disinfection

Frequency: After Each Patient

3 STEPS TO DISINFECT


STEP 1

Using an EPA* or DIN** registered disinfectant wipe, gently wipe the stethoscope



STEP 2

The stethoscope must remain wet with the disinfectant solution for the duration of the contact time as per the label directions for use



STEP 3

After the contact time has been achieved, leave to air dry or wipe dry with a clean cloth



*United States Environmental Protection Agency

**Health Canada Drug Identification Number